

GOVERNMENT OF INDIA  
BCG VACCINE LABORATORY  
GUINDY, CHENNAI – 32

STORES : 044-22500172  
SECURITY : 044-22501740  
FAX : 91-44-22501047

File No. D-21007/04/2019-20 (S)

Dated 12.06.2019


CORRIGENDUM

\*\*\*\*\*

The following amendment to this office Tender Enquiry No. D-21007/04/2019-20 (S) dated 30.05.2019 for supply of Digital Analytical Weighing Balance with Printer are hereby authorized:

Specification for Digital Analytical Weighing Balance	FOR	READ
Readability	0.1 mg	0.01 mg
Weighing capacity	320 g	120 gm
Leveling	Automatic	Level sensor alarm & guidance
Adjustment	Internal	No change
Display	Digital, Black and White	Digital, Color
Body	Vendor to specify	MOC of Pan-SS 316
Connection	Battery-operated, plug in adapter	Plug in adapter
Other additions		Password protection
		Printer attachment – GLP Compliant print out
		Calibration – Internal / Automatic
Date & Time of Tender Closing	2.00 PM on 26.06.19	2.00PM on 04.07.19
Date & Time of Tender Opening	3.00PM on 26.06.19	3.00PM on 04.07.19

All other terms and conditions remain unaltered.

  
VETERINARIAN  
FOR DIRECTOR

To

As per list enclosed.