

GOVERNMENT OF INDIA  
BCG VACCINE LABORATORY  
GUINDY, CHENNAI – 32

STORES : 044-22500172  
SECURITY : 044-22501740  
FAX : 91-44-22501047

File No.D-21007/3/2019-20(S)

Dated:28.05.2019

As per list enclosed.

Dear Sir,

**Sub.:** Quotation for the supply of pH meter and pH probe with printer for testing of Low conductivity water as well as prepared media – reg.

\*\*\*\*\*

Please Forward Your "**Sealed quotation**" superscribed on the envelope for supply of the following materials to reach this office **immediately / on or before 26.06.2019.**

SI.No.	DESCRIPTION OF ITEMS	PROBABLE REQUIREMENT
1.	pH meter with printer	2 Nos.
2.	Low Conductivity pH Probe	1 No.

**Specifications for SI. No. 1 (pH meter with Printer)**  
**(compatible with low conductivity pH probe)**

Channel:Single channel  
pH Range:-2 to 20 pH  
pH Resolution: 0.001; 0.01; 0.1  
pH Accuracy: +or -0.002  
Calibration Points: Up to 5 points with Auto-buffer recognition  
mV Resolution: 0.1; 1  
mV Accuracy: + or - 0.1  
Temperature Range:0<sup>0</sup> C to 100<sup>0</sup> C  
Temperature Resolution: 0.1<sup>0</sup> C  
Temperature Accuracy: + or - 0.1<sup>0</sup> C  
Probe Compatibility: To be compatible with low conductivity pH probe  
Printer : Yes

**Specifications for SI. No. 2 (Low Conductivity pH probe):**

Reference: Single, Ag/Agcl  
Junction: Ceramic, triple  
Connector Type: To be compatible with pH meter  
Electrolyte: KCL 3.5M + Agcl  
Maximum Pressure : 0.1 bar

: 02 :

pH Range: 0 to 12 pH at 0 to 70° C

Tip Shape: Spherical(dia: 9.5mm)

Temperature Sensor: Yes

Amplifier: Yes

Body Material : Glass

Recommended use : To be used for water – for – injection, purified water and raw water

**Pre-bid meeting will be conducted at 11.00 AM on 04.06.2019**

**Date & Time of Closing of Tender : 02.00 p.m. on 26.06.2019**

**Date & Time of Opening of Tender : 03.00 p.m. on 26.06.2019**

**IMPORTANT NOTE:**

1. Quotation should be valid for 90 days.
2. The envelope should be superscribed with "**Tender No. D-21007/03/2019-20(S) due for opening on 26.06.2019**".
3. Rate inclusive of all taxes with break up should be given in sealed cover superscribed with enquiry number and date.
4. Payment will be made against submission of Tax Invoice (in triplicate).
5. Your quotation should not be pre-conditional. Pre-conditional quotations will be summarily rejected.
6. Rates should be quoted on FOR at BCGVL, Chennai -32.
7. Rates should be quoted in INDIAN rupees only.

Yours faithfully,

  
ADMINISTRATIVE OFFICER  
For DIRECTOR, BCGVL